

NB 2163

EUTYPE EXAMINATION CERTIFICATE

Certificate No: 2163-PPE-1859

Respiratory protective devices, filtering half masks to protect against particles manufactured for

TULPAR SAĞLIK ÜRÜNLERİ İMALAT SAN.VE LTD. ŞTİ.

Konutkent mah. 3035 Cadde No:74/B28 Çankaya, Ankara TURKEY

by the following manufacturer

Ece Dermokozmetik Ltd. Şti.

Tevfikbey Mah. Şehit Erol Olçok Cad. No:19 İç Kapı No:1 Küçükçekmece, İstanbul TURKEY

are tested and evaluated according to

EN 149:2001 + A1:2009 Respiratory Protective Devices -Filtering Half Masks to Protect Against Particles -Requirements, Testing, Marking

Based on the type examination conducted with the evaluation of test reports, technical file according to Personal Protective Equipment Regulation (EU) 2016/425 Annex 5, it is approved that the product meets the requirements of the regulation.

Product Definition

Single use particle filtering half mask for protection against solid and liquid aerosols, is a fish type, 5 layered, without valve, ear straps and adjustable nose bar.

Brand Name: DNA Model: 2972FM Classification: FFP3 NR

For more details, refer technical evaluation report provided to the manufacturer, dated 28.12.2020 and number 2163-KKD-1859.

Here by the manufacturer is allowed to use notified body number (2163) and can fix CE mark, as shown below, on the Category III product models given above, with;

- Issuing an appropriate EU Declaration of Conformity according to Personal Protective Equipment Regulation (EU) 2016/425 Annex 9.
- Ongoing successful performance in fulfilment of the requirements set out in Personal Protective Equipment Regulation (EU) 2016/425 and harmonised standards, ensured by assessments based on Annex 7 (Module C2) or Annex 8 (Module D) of the regulation no later than 1 year from the beginning of serial production

This certificate is initially issued on 28/12/2020 and will be valid for 5 years, if there is no change in the relevant harmonised standard affecting the essential health and safety requirements.



UNIVERSAL CERTIFICATION Director